



How should I reduce paravalvular leak for my TAVI patients ?

Prof. Dave Smith – Swansea, United Kingdom

Dr. Anthony Camuglia – Brisbane, Australia

Dr. Michael Kang-Yin Lee – Kowloon, Hong Kong

Question 1: Are there any anatomies or clinical situation where the valve is not suitable, or you would prefer a different valve choice?

Currently Navitor is not suitable to treat larger annular anatomies (perimeter over 85mm, area over 573mm²). Also, Navitor has little data for use in bicuspid anatomy and so should be used on an individual basis after heart team review.

PCR

PCRONline.com